













RAYCAST® HIGH PRECISION ACCESSORIES HEAD SUPPORTS
LOW DENSITY FOR PUSH-PIN HARDWARE
HIGH PRECISION LD HEAD SUPPORTS FOR MULTIPOINT
BASE PLATES WITH PUSH-PINS

Models	Pictures	Dimensions
Model 1 with lateral neck flaps Article No. : 32414		
Model 2 without lateral neck flaps Article No. : 32415		
Model 3 with lateral neck flaps Article No. : 32412		
Model 4 without lateral neck flaps Article No. : 32413		
Model 5 with lateral neck flaps Article No. : 32410		
Model 6 without lateral neck flaps Article No. : 32411		
Article No.: 32409	Set of all the above	

A. GENERAL PRODUCT INFORMATION

The products referred to in these instructions are medical devices used for patient positioning and immobilisation in radiation therapy. These low density head supports can be used during both the simulation and treatment stage.

To attain an optimal result, it is recommended to use this product in combination with Orfit immobilisation products.


B. PRODUCT DESCRIPTION

These LOW DENSITY HEAD SUPPORTS FOR MULTIPOINT BASE PLATES WITH PUSH PINS are used in combination with base plates that have a rectangular cut-out for the head supports and with EFFICAST® PRE-CUTS. Together with these parts they form a reproducible patient positioning and immobilisation system. The head supports offer the opportunity to position the patient's head in a stable, reproducible way. Information on the EFFICAST PRE-CUTS can be found in the respective 'instructions for use' and on www.orfit.com.

C. PRECAUTIONS FOR USE

These devices can only be used on the above described base plates.

The head supports are securely positioned on the base plates by means of a positioning block made of green foam on the bottom of the head support that fits the recess in the base plates.



Always verify that the devices are positioned correctly on these base plates.

D. STORAGE

Always store the product in a safe place to prevent them from getting damaged. Do not put heavy objects on the devices and prevent hard objects from falling onto it, because these will cause permanent damage.

Store the devices between +10°C (50°F) and 40°C (122°F).

E. PROPERTIES

E.1. Basic Physical Properties

The following physical properties apply to these devices:

Dimensions:	L 230 - 257 mm
	W 108 - 120 mm
	H 82 - 92 mm
Weight:	0,05 - 0,07 kg

E.2. Material Properties

The LD head supports are made of PE foam covered with a PU coating and have a good dimensional stability.

This product is made entirely of electrically non-conductive, non-metallic and non-magnetic materials and is MR safe.



E.3 Dosimetric Properties

These head supports are made of low density foam that has the following dosimetric properties per cm of material through which the beam passes:

Attenuation at 6 MV: 0.15 % per cm.
Attenuation at 15 MV: 0.05 % per cm.

Skin build-up: 0.39 mm H₂O equivalent per cm.

Note: Use these numbers as a guidance only. Perform the measurements again in your department to verify these results.

F. MAINTENANCE AND WASTE MANAGEMENT

These products can be cleaned and disinfected by means of an isopropanol based disinfectant, applied with a soft cloth. If unsure about the cleaning fluid, do not use. **Never use aerosol sprays, corrosive cleaning agents, solvents or abrasive detergents. Do not soak the cushions.** Further cleaning instructions can be found in the [Orfit Cleaning guidelines](#). Periodic checks of the product should be done to insure the parts are not worn and require repair or replacement. **Do not attempt to make repairs yourself.** Contact your distributor if there are any questions or concerns.

The products can be disposed of with household waste.

G. ADDITIONAL INFORMATION

For additional information such as distributor contact information, product brochures, Safety Data Sheets and regulatory information, please visit our website www.orfit.com.

Note:

It is prohibited to make alterations to this text without prior approval from Orfit Industries. RAYCAST® and EFFICAST® are registered trademarks of Orfit Industries.



ORFIT INDUSTRIES
Vosveld 9A | B-2110 Wijnegem | Belgium
T (+32) (0)3 326 20 26
welcome@orfit.com

www.orfit.com



Ref. No. 31945
VERSION 10
LAST UPDATE: 01/04/2021
REVISION DATE: 01/04/2023